


## FULL VIEW SIGHT FLOW SIGHT TUBE REPAIR KIT ORDERING OPTIONS:



**ORDER CODE**      **CODE DESCRIPTION**

**CONNECTION TYPE:**  
 FVSF-F    FLANGED  
 FVSF-T    THREADED

**CONNECTION SIZE:**  
 0.50      1/2"  
 0.75      3/4"  
 1.00      1"  
 1.25      1-1/4" (FVSF-T ONLY)  
 1.50      1-1/2"  
 2.00      2"  
 2.50      2-1/2" (FVSF-F ONLY)  
 3.00      3" (FVSF-F ONLY)  
 4.00      4" (FVSF-F ONLY)  
 6.00      6" (FVSF-F ONLY)

**SEAL MATERIAL:**  
 B          BUNA-N  
 E          ETHYLENE PROPYLENE  
 N          NEOPRENE  
 V          FLUOROCARBON  
 TN        TEFLON ENVELOPE WITH NEOPRENE FILLER  
 TS        TEFLON ENVELOPE WITH SILICONE FILLER  
 TV        TEFLON ENVELOPE WITH FLUOROCARBON FILLER

**LENGTH IN INCHES:**  
 BLANK    LEAVE BLANK FOR STANDARD LENGTH  
 X.XXX    SPECIAL LENGTH TO BE A THREE PLACE DECIMAL

**SIGHT TUBE REPAIR KIT:**  
 RK

**EXAMPLE:** A SIGHT TUBE REPAIR KIT FOR A MODEL FVSF-F-0.50-B-CS ARMORED SIGHT FLOW GAUGE WITH 1/2" FLANGED CONNECTIONS, BUNA-N SEALS AND CARBON STEEL CONSTRUCTION SHOULD BE ORDERED AS FVSF-F-0.50-B-RK.

| REV. | DATE     | BY  | CHKD. | ECN    | EXPLANATION:                           | KENCO ENGINEERING COMPANY<br>TULSA, OKLAHOMA                 |
|------|----------|-----|-------|--------|--|--|
| A    | 09-20-10 | PAS | KI    | 000003 | SEE SHEET 1 OF 2                       | GUIDE, ORDERING FOR KENCO<br>FULL VIEW SIGHT FLOW INDICATORS |
| B    | 12-29-11 | PAS | RDM   | 000003 | SEE SHEET 1 OF 2                       |  |
| C    | 03-15-13 | KI  | DS    | 000193 | SEE SHEET 1 OF 2                       |  |
| D    | 09-05-13 | RDM | DM    | N/A    | ADDED INSTRUCTIONS FOR SPECIAL LENGTHS |  |
|      |          |     |       |        |  | DATE: 09-20-10    SHEET: 2 OF 2    DRAWING NUMBER            |
|      |          |     |       |        |  | DRN. BY: PAS      CHKD. BY: KI      REVISION: D      C-61322 |
|      |          |     |       |        |  | SCALE: NONE  |